

## **Certificate of Analysis**

COA No: CA\_BSM-0007

Version: 08

## SensiFAST™ SYBR® & Fluorescein Kit

For research or further manufacturing use only

Catalog No:	BIO-96020
Lot No:	SF585-B112800
Storage Conditions:	-20°C
Component Lot No:	SFSF-222111A
Expiry date:	December 2024

## **Quality Control Parameters**

Analysis	Specification	Result
Functional	Quantitative PCR analysis amplifying 6 genes from a dilution series of mouse cDNA under standard conditions. Cq and melting profiles must be consistent for the test and reference sample with $\pm0.5$ Cq variance.	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with control sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at $37^{\circ}$ C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection $2.5 \times 10^{-3}$ U DNase I.	Passed

QA / QC Representative:

Andrew Galeeba-M

United Kingdom USA

Tel: +44 (0)20 8830 5300 Fax: +44 (0)20 8452 2822 Tel: +1 901.382.8716 Fax: +1 901.382.0027 <u>Germany</u>

Tel: +49 (0)3371 60222 00 Fax: +49 (0)3371 60222 01 <u>Australia</u>

Tel: +61 (0)2 9209 4180 Fax: +61 (0)2 9209 4763

Date: 30<sup>th</sup> November 2022